

## DETERMINATION OF BLEEDING OF CONCRETE AS 1012.6

This standard describes the procedures for determining the relative quantity of mixing water that will bleed from a sample of freshly mixed concrete and applies only to concrete in which the maximum aggregate size is 40mm or less and provides for compaction of the sample by hand tamping or vibration.



### CL 10094 BLEEDING CONTAINER

The Civilab Bleeding Container is manufactured from heavy gauge steel plate and reinforced to maintain its shape under rough usage. It has an inside diameter of 250mm and a height of 280mm. It has been machined top and bottom to make it water tight. A sealed lid is provided. The container is fitted with carrying handles.



CL10094

### CL10095 BLEEDING CONTAINER SPLIT TYPE

The Civilab Split Type Bleeding Container complies to AS-1012.6.1999. It is constructed of 6mm sheet rolled steel with air tight seal. It has 2 lug and bolt arrangements 55mm from top and bottom. 250mm diameter x 280mm high.

#### ACCESSORIES:-

CL 40409	SUCTION DEVICE
CL 90571	MEASURING CYLINDER 25ml
CL 90573	MEASURING CYLINDER 100ml
CL 10098	ELECTRONIC BALANCE 30kg x 10gms
CL 10012	TAMPING ROD
CL 10014	SCOOP
CL 10097	RUBBER MALLET
CL 10150	VIBRATOR INTERNAL
CL 10155	VIBRATOR EXTERNAL